
Project BREATHE – Prototype Respirator Evaluation Utilizing Newly Proposed Respirator Test Criteria

Jung-Hyun Kim, Raymond J. Roberge*, Ronald E. Shaffer, Ziqing Zhuang, Jeffrey B. Powell, Michael Bergman, and Andrew J. Palmiero

National Institute for Occupational Safety and Health, National Personal Protective Technology Laboratory, Pittsburgh, Pennsylvania 15236

* Corresponding author and E-mail: dtn0@cdc.gov

ABSTRACT

Machine and human subject testing of four prototype filtering facepiece respirators (FFR) and two commercial FFR was carried out utilizing recently proposed respirator test criteria that address healthcare worker-identified comfort and tolerance issues. Overall, two FFR (one prototype, one commercial model) were able to pass all eight criteria and three FFR (two prototypes, one commercial model) were able to pass seven of eight criteria. One prototype FFR was not tested against the criteria due to an inability to obtain satisfactory results on human subject quantitative respirator fit testing. Future studies, testing different models and styles of FFR against the proposed criteria, will be required to gauge the overall utility and effectiveness of the criteria in determining FFR comfort and tolerance issues that may impact user compliance and, by extension, protection.

Keywords: filtering facepiece respirators, prototypes, proposed respirator test criteria.

INTRODUCTION

Project BREATHE (Better Respiratory Equipment utilizing Advanced Technology for Healthcare Employees) was a joint undertaking of the National Personal Protective Technology Laboratory (NPPTL) of the National Institute for Occupational Safety and Health (NIOSH) and the Veterans Health Administration (VHA) that sought to develop a respirator(s) designed specifically for healthcare workers (HCW) and provisionally termed the “B95 Respirator” (Radonovich *et al*, 2009; Gosch *et al*, 2013). Seeking outside collaborators, this undertaking was announced by Department of Veterans Affairs in the Federal Register (Federal Register, 2009) with the result that two U.S. respirator manufacturers (Company A, Company B) responded and agreed to participate in this endeavor. Manufacturer-supplied prototype respirators were evaluated with human subject testing (physiological, subjective) and machine testing (filter penetration, breathing resistance) and graded on a pass / fail basis against recently-established metrics for respirator evaluation (Shaffer *et al*, 2014). While previous publications communicated the vision for Project BREATHE (Gosch *et al*, 2013), establishing pass-fail criteria and test methods (Shaffer *et al*, 2014), the purpose of this study was to apply these concepts to actual products and possible future products. This report summarizes the findings of this testing.

MATERIALS AND METHODS

Prototype Respirators

Each manufacturer supplied two models of prototype respirators, each with filters of the N95 class and none of which was equipped with an exhalation valve. The prototypes provided were concept level working prototypes and did not represent final design lockdown of a product nor the refinement that would be expected of a mass production sample.

Company A Prototype Respirator 1 (A-1) - a disposable rigid, cup-shaped filtering facepiece respirator model available in small and standard sizes with two non-adjustable elasticized straps and a pliable plastic nose bar.

Company A Prototype Respirator 2 (A-2) - a disposable pliable, V-shaped, pleated filtering facepiece respirator model available in small and standard sizes with two non-adjustable elasticized straps and a pliable plastic nose bar.

Company B Prototype Respirator 1 (B-1) – a reusable filtering facepiece respirator/elastomeric respirator hybrid available in small, medium and large sizes, comprised of a pliable, opaque silicone facemask with a centrally located, vertically-positioned, parabolic-shaped, replaceable filter and filter housing with a single piece elasticized harness.

Company B Prototype Respirator 2 (B-2) – a filtering facepiece respirator, available in small and medium / large sizes, with a curved, horizontally-positioned, oblong-shaped, plastic frame that houses a central filter panel, to which is attached a chin panel and a less permeable nasal panel with adjustable aluminum nasal bar, and an adjustable single piece strap.

Other Respirator Testing – two commercially-available surgical N95 filtering facepiece respirator models (3M 1860, 3M 1870; 3M Company, St. Paul, MN), that have received the most previous research attention for this class of respirators (Shaffer *et al*, 2014; Shaffer and Jansen, 2015) and are commonly available in healthcare facilities (Wizner *et al*, 2016), were tested with the same test criteria. Both models were tested with the subjects used in Company A prototype trials (n=21). In addition, model 3M 1870 was tested with the subjects used in Company B prototype trials (n=20).

Machine Testing

Each model of the four prototype respirators was tested individually five times for breathing resistance and for filter penetration of nebulized sodium chloride solution with the Certi Test 8130 filter tester (TSI, Shoreview, MN). NIOSH pass criteria for N95 category respirator filters requires filter penetration $\leq 5\%$ and inhalation resistance ≤ 35 mm H₂O pressure tested at a continuous flow rate of 85 L/min (NPPTL, 2012). Filter penetration is not included in the currently proposed respirator test criteria, but was performed in the current study to ensure the requisite protective qualities of the prototype respirators filters. Five respirator samples of each prototype underwent 10 donnings followed by quantitative respirator fit testing ($n=50$ fit tests for each prototype respirator model) on the Static Advanced Headform (Hanson Robotics, Inc., Plano, Texas) that utilizes Frubber™, a fluid-filled cellular matrix composed of an elastomer that simulates the physics of human facial living soft tissues as the simulant skin covering. Respirator fit testing on the Static Advanced Headform has previously been shown to correlate with human results (Bergman *et al*, 2015). Passage of a NIOSH respirator quantitative fit test for N95 class respirators is based on a minimum required fit factor of 100 that is calculated as the ratio of measured ambient particles to respirator deadspace particles (OSHA, Appendix A, 2003).

Human Subject Quantitative Respirator Fit Testing

Eighteen men (anthropometrics, 32.3 ± 14.4 yrs, height 179.4 ± 8.9 cm, weight 81.5 ± 17.5 kg, BMI 25.5 ± 5.0 kg/m²) and 17 women (anthropometrics, 32.9 ± 12.4 yrs, height 165.3 ± 6.2 cm, weight 77.2 ± 21.4 kg, BMI 28.4 ± 8.2 kg/m²), representative of the cells in the NIOSH respirator fit test panel (Zhuang and Bradtmiller, 2007), underwent quantitative fit testing (OSHA, Appendix A, 2003) on each of the four prototypes. Passage of an OSHA respirator quantitative fit test for N95 class respirators is based on a minimum required fit factor of 100 that is calculated as the ratio of measured ambient particles to particles within the deadspace of the respirator (OSHA, Appendix A, 2003). Respirators were tested as a “family” and up to 2 donnings per size were permitted as indicated in the proposed criteria (Shaffer *et al*, 2014). A researcher selected a sample of the size most likely to provide a good fit. If the subject failed 2 donnings, another size was selected and tested. Upon passage of a fit test, no further sizes of the respirator were tested. Similar methods were employed elsewhere (Zhuang *et al*, 2017) in development of proposed respirator fit capability (RFC) test criteria.

Human Physiological/Comfort Studies – Subject Anthropometrics

Subject Test Group 1 (Company A trials) - Twenty-one healthy, non-smoking subjects (12 men, 9 women) participated in the study exercise trials. Subject anthropometrics for men were age 22.6 ± 2.7 yrs, height 180.8 ± 7.9 cm, weight 84.1 ± 16.5 kg, and Body Mass Index (BMI) 25.7 ± 4.8 kg/m². Anthropometric values for women were age 23.7 ± 3.3 yrs, height 166.4 ± 4.0 cm, weight 65.6 ± 6.5 kg, and BMI 23.7 ± 2.6 kg/m². Each of these subjects also underwent testing with the 3M 1860 and 1870 models.

Subject Test Group 2 (Company B trials) - Twenty healthy, non-smoking subjects (10 men, 10 women) participated in the study. Subject anthropometrics for men were age 23.1 ± 1.6 yrs, height 181.7 ± 7.2 cm, weight 78.9 ± 8.6 kg, and BMI 23.9 ± 2.9 kg/m². For women, anthropometrics were age 22.0 ± 2.5 yrs, height 165.5 ± 3.0 cm, weight 62.5 ± 5.2 kg, and BMI 22.8 ± 2.0 kg/m². Each of these subjects also underwent testing with the 3M 1870 model.

The exercise trials were conducted in a physiology laboratory with mean ambient temperature $20.3 \pm 1.2^\circ\text{C}$, relative humidity $32.6 \pm 14.4\%$, and barometric pressure 739.1 ± 2.5 mm Hg during the study period. These ambient conditions are similar to those encountered by HCW, the single largest group of respirator users (Smith and Rea, 1977). On the initial day of testing, subjects underwent a screening history and physical examination by a licensed physician. Subjects were dressed in athletic shorts or

pants, tee shirt and athletic shoes during exercise testing. The study was approved by the NIOSH Human Subjects Review Board, and all subjects provided oral and written informed consent.

Subject Instrumentation

Subjects were instrumented for continuous physiological monitoring during the exercise trials. The heart rate (HR), oxygen saturation (SpO₂) and transcutaneous carbon dioxide (tcpCO₂) were continuously monitored with the Tosca 500™ (Radiometer, Copenhagen, DM), an earlobe-mounted combination pulse oximeter and heated, Severinghaus-type CO₂ sensor that was adhesively attached to the left earlobe (Roberge *et al*, 2010). The respiratory rate (RR) was monitored with the Zephyr Bioharness™ (Medtronic, Minneapolis, MN, US), an elasticized chest strap utilizing a proprietary embedded capacitive sensor to evaluate chest expansion and contraction (Kim *et al*, 2012). Tympanic membrane temperatures (T_{tymp}) were measured by a single investigator (research physician) at baseline and pre-and-post exercise trials from the right ear with a Welch/Allyn Pro 4000 infrared tympanic thermometer (Braun GmbH, Kronberg, FRG). Respirator microclimate (deadspace) temperature and relative humidity were continuously monitored with the iButton (Maxim, San Jose, CA, US), a small (16 mm x 6 mm) wireless sensor that was adhesively attached to the inner surface of the N95 FFRs midway between the respirator center and the edge of its right upper quadrant (Roberge *et al*, 2012a). The temperature of the facial skin covered by the respirators was measured with the VitalSense (Philips Respironics, Bend, OR, US) wireless dermal sensor patch (Roberge *et al*, 2012b).

Exercise Protocol

The study trials consisted of walking on a treadmill at 5.6 km/h (0° incline) wearing one of five respirators (A-1, A-2, B-1, 3M 1860, 3M 1870) for 1 hr each. Respirator B-2 was not tested because it did not pass fit testing in sufficient numbers to meet the respirator testing criteria (Shaffer *et al*, 2014). No more than two trials were conducted in a single day with a minimum one-half hour respite between any trials. For trials, the subjects donned a randomized respirator and performed a user seal check (OSHA, Appendix B-1, 2003) immediately prior to the treadmill session. Subjective ratings of exertion and thermal comfort during the trial sessions were assessed at timed intervals (0, 15, 30, 45 and 60 minutes) with the Borg Perceived Exertion Scale (a 15-grade scale ranging from “no exertion at all” to “maximal exertion”) (Borg G, 1990) and the Frank Comfort Scale (a 10-point scale ranging from “the coldest you have ever been” to “the hottest you have ever been”) (Frank *et al*, 1999). Additionally, during the trials, two 7-point respiratory scales for perceived inspiratory effort and perceived expiratory effort (ranging from “not noticeable” to “intolerable”), and one 7-point respiratory scale for overall breathing discomfort (ranging from “no discomfort” to “intolerable discomfort”) were administered (Antunano *et al*, 1993). Between sessions, subjects were seated and allowed to drink bottled water or a sports drink *ad lib*.

Data Analysis

Study data values for physiological and subjective variables during exercise trials were calculated by differences between mean baseline (0min) and mean 60min values. Prototype respirators and the two commercially available respirator models were not compared to one another for the study, but were individually evaluated by comparison with eight previously-cited proposed respirator test criteria (Shaffer *et al*, 2014) on a pass/fail basis (see Table I for criteria). Machine test study data for filter penetration were evaluated against NIOSH established testing norms (NPPTL, 2012; Shaffer *et al*, 2014).

RESULTS

The mean physiological and subjective measurement data during exercise trials for each of the three prototype respirators and the 3M model 1860 and 1870 respirators tested are dichotomized by subject group (Test Groups 1 and 2) and presented in Table 2. As presented in Table 1, Respirator A-2 and 3M model 1870 were able to pass 8/8 criteria, whereas A-1, 3M model 1860 and B-1 were able to pass 7/8 respirator test criteria (Shaffer *et al*, 2014). Respirator A-1 and 3M model 1860 each failed to achieve a Geometric Mean Manikin Fit Factor of ≥ 100 on 100% of the five samples ($n=50$ donnings per respirator model) tested with the Advanced Static Headform testing. Respirator B-1 failed the breathing resistance criteria with a mean filter resistance of 13 mm H₂O pressure. Respirators A-1, A-2, B-2, 3M 1860, and 3M 1870 had filter resistances (mm H₂O) of 8.5, 3.7, 5.6, 8.9, and 6.3 respectively. Respirator B-2 achieved a pass rate of 30% on quantitative respirator fit testing (Table 1). All tested respirators were able to achieve NIOSH passing results on filter penetration testing. The four prototype respirators all had filter penetration values less than 1%.

Table 1. Pass (P) / Fail (F) Results of Prototype Respirator Testing Utilizing Recently Proposed Respirator Test Criteria (Shaffer *et al*, 2014)

Variables Tested	RESPIRATORS TESTED					
	A-1	3M 1860	A-2	3M 1870	B-1	B-2
Breathing Resistance ≤ 10 mm H ₂ O	P	P	P	P	F	N/A
tcPCO ₂ level increase < 4 mm Hg *	P	P	P	P	P	N/A
SpO ₂ decrease $\leq 1\%$ *	P	P	P	P	P	N/A
Moisture retention $\leq 4\%$ *	P	P	P	P	P	N/A
Facial skin microclimate heat $\leq 2.5^{\circ}\text{C}$ increase over baseline*	P	P	P	P	P	N/A
Respirator microclimate heat $\leq 2.5^{\circ}\text{C}$ increase over baseline*	P	P	P	P	P	N/A
Reuse/Gauging Fit; manikin geometric mean fit factor ≥ 100 on 5/5 samples§	F	F	P	P	P	F
Pass rate of $\geq 74\%$ on OSHA quantitative fit testing	P	P	P	P	P	F

* Over 1 hr.

§ Tested with the Advanced Static Headform.

Table II. Subjects' Mean Physiological and Subjective Measurement Data ($\pm$ standard deviation) during Exercise Trials Wearing Three Prototype Respirators and Two Models of Commercially-Available Respirators

Study Variable		Subject Test Group 1				Subject Test Group 2	
		A-1	3M 1860	A-2	3M 1870	B-1	3M 1870
SpO ₂ %	0min	98.4 $\pm$ 1.0	96.8 $\pm$ 6.2	98.3 $\pm$ 0.9	98.1 $\pm$ 1.0	98.8 $\pm$ 0.8	98.4 $\pm$ 0.8
	60min	98.4 $\pm$ 1.0	97.8 $\pm$ 2.2	98.2 $\pm$ 0.8	98.3 $\pm$ 0.9	98.4 $\pm$ 0.8	98.4 $\pm$ 0.9
tcpCO ₂ mm Hg	0min	40.2 $\pm$ 4.8	38.0 $\pm$ 4.5	38.9 $\pm$ 4.3	38.6 $\pm$ 4.2	38.0 $\pm$ 3.0	37.9 $\pm$ 3.2
	60min	40.3 $\pm$ 4.8	40.3 $\pm$ 4.8	39.2 $\pm$ 4.7	39.3 $\pm$ 4.4	40.4 $\pm$ 3.4	40.1 $\pm$ 3.8
HR	0min	100.3 $\pm$ 17.6	96.9 $\pm$ 18.0	93.5 $\pm$ 19.8	94.9 $\pm$ 19.2	79.6 $\pm$ 15.9	83.4 $\pm$ 13.8
	60min	111.1 $\pm$ 20.4	109.3 $\pm$ 21.7	107.9 $\pm$ 21.9	106.2 $\pm$ 23.3	115.7 $\pm$ 19.7	113.2 $\pm$ 20.0
RR	0min	20.9 $\pm$ 5.1	21.0 $\pm$ 4.9	21.0 $\pm$ 4.7	21.1 $\pm$ 4.4	17.7 $\pm$ 4.6	16.1 $\pm$ 3.6
	60min	32.0 $\pm$ 9.8	29.1 $\pm$ 11.9	28.3 $\pm$ 5.7	30.5 $\pm$ 6.5	26.9 $\pm$ 4.5	25.7 $\pm$ 5.3
T _{tympanic} °C	0min	36.4 $\pm$ 1.0	36.4 $\pm$ 0.6	36.7 $\pm$ 0.2	36.5 $\pm$ 0.3	36.6 $\pm$ 0.3	36.4 $\pm$ 0.3
	60min	36.6 $\pm$ 0.5	36.6 $\pm$ 0.4	36.7 $\pm$ 0.3	36.6 $\pm$ 0.3	36.9 $\pm$ 0.4	36.9 $\pm$ 0.4
Inspiratory Effort	0min	1.7 $\pm$ 0.4	1.6 $\pm$ 0.4	1.4 $\pm$ 0.6	1.5 $\pm$ 0.5	1.2 $\pm$ 0.4	1.1 $\pm$ 0.2
	60min	2.3 $\pm$ 0.9	2.2 $\pm$ 0.8	2.3 $\pm$ 1.0	2.1 $\pm$ 0.9	2.3 $\pm$ 0.8	2.4 $\pm$ 0.7
Expiratory Effort	0min	1.7 $\pm$ 0.5	1.6 $\pm$ 0.4	1.5 $\pm$ 0.5	1.7 $\pm$ 0.5	1.2 $\pm$ 0.5	1.0 $\pm$ 0.0
	60min	2.4 $\pm$ 0.8	2.2 $\pm$ 0.7	2.3 $\pm$ 0.9	2.2 $\pm$ 0.7	2.4 $\pm$ 0.8	2.4 $\pm$ 0.7
Overall Breathing Difficulty	0min	1.5 $\pm$ 0.5	1.5 $\pm$ 0.5	1.4 $\pm$ 0.5	1.4 $\pm$ 0.5	1.2 $\pm$ 0.5	1.1 $\pm$ 0.2
	60min	2.3 $\pm$ 0.9	2.4 $\pm$ 0.8	2.5 $\pm$ 1.3	2.1 $\pm$ 0.8	2.3 $\pm$ 0.8	2.3 $\pm$ 0.6
Borg Scale	0min	8.5 $\pm$ 1.8	9.2 $\pm$ 2.3	8.5 $\pm$ 1.8	9.0 $\pm$ 2.1	5.9 $\pm$ 1.2	6.2 $\pm$ 0.4
	60min	11.0 $\pm$ 2.9	11.0 $\pm$ 2.8	11.0 $\pm$ 2.5	10.9 $\pm$ 2.7	9.5 $\pm$ 2.6	10.4 $\pm$ 2.0
Frank Scale	0min	4.5 $\pm$ 0.7	4.8 $\pm$ 0.7	4.7 $\pm$ 0.8	4.8 $\pm$ 0.6	4.3 $\pm$ 1.0	4.3 $\pm$ 0.7
	60min	6.3 $\pm$ 0.8	6.4 $\pm$ 0.8	8.6 $\pm$ 0.5	6.2 $\pm$ 0.7	6.2 $\pm$ 1.2	6.5 $\pm$ 0.7

DISCUSSION

Respiratory protective devices were initially developed for industrial workers, not with HCW in mind (Gosch *et al*, 2013). There are a number of features of industrial work (dusty environments, heavy physical workloads, extreme ambient conditions, etc.) that are not generally present in the typical healthcare setting, suggesting that the respiratory protection needs of HCW may be different and better served with respiratory protective devices tailored to their environment. Project BREATHE (Radonovich *et al*, 2009) was an effort to partner the U.S. federal government with manufacturers in developing a HCW-specific respirator(s) that could be tested against metrics that have been identified from previously-completed respirator research projects (Shaffer *et al*, 2014).

The current study is a first attempt at testing prototype respirators, developed with the HCW market in mind, utilizing pass/fail criteria from recently established respirator test criteria (Shaffer *et al*, 2014). The study data suggest that current respirator manufacturing technology is able to develop respiratory protective devices that can meet the proposed criteria for a HCW-specific respirator(s). This is important in that the recently proposed criteria were developed to evaluate parameters thought to influence user comfort and thereby compliance, while not sacrificing protection. Among the three prototypes undergoing human subject testing, the A-2, A-1 and B-1 prototypes were able to achieve a passing score on 8/8, 7/8 and 7/8 respirator test criteria, respectively, and A-2 ultimately went on to be commercialized. These three prototype models passed the criteria for fit testing with human quantitative fit test results equal to, or better than, those noted for respiratory protective devices currently used in the healthcare setting (Wilkinson *et al*, 2010). Respirators A-1 and A-2 had pass rates of 83%, while all 35 subjects (100%) passed with the B-1. The similar higher HRs and associated lower RRs noted for B-1 and 3M 1870 at 60 min, compared with the other respirators tested, suggest a greater effort at the work of breathing that could be related to filter resistance. However, with the exception of B-1, all the tested respirators had filter resistances that were <10 mm H₂O resistance, a level that has previously been shown to result in HRs and RRs that are not significantly different from controls without a respirator (Roberge *et al*, 2013). Therefore, these differences in RR and HR responses may represent non-respirator related variables such as physical conditioning differences between the two study groups given that, at low and moderate work rates, workload intensity has a greater physiological impact on HR and RR than the magnitude of inspiratory resistance (Antunano *et al*, 1993).

Respirator prototype testing using the newly proposed criteria may allow for refinement of some respirator features such as improved fit and decreased facial respirator microenvironment temperature elevations, whereas other endeavors, such as attempting further decreases in filter respirator resistance, may not be beneficial or noticeable as evidenced by the similar physiological (e.g., RR) and subjective responses (e.g., inspiratory effort, expiratory effort, overall breathing difficulty) shared by the three prototypes that underwent human testing (Table II). One limitation of the current study is that all of the machine and human subject testing was conducted in a laboratory setting. Confirmation of study findings is still needed in large clinical trials. Field testing of the current study's three prototype respirators in VHA healthcare settings has recently been accomplished and the forthcoming results of that testing should provide further important information (Radonovich *et al*, 2009). Another limitation of this study is that data from only 8 of the 10 proposed B95 tests could be used. Two of the criteria related to "air exchange" (i.e., average inhaled carbon dioxide and oxygen) relied upon NIOSH's Automated Breathing and Metabolic Simulator (ABMS). Unfortunately, problems occurred during testing that resulted in the data being unusable. However, human subject test data also addressing "air exchange", namely transcutaneous CO₂ and O₂ saturation, were found to be consistent when comparing the repeat 3M 1870 testing (Table 2), suggesting that ABMS testing may not be needed going forward. The newly proposed, metrics-based respirator test criteria may allow for a more robust evaluation of future respirator feature modifications and improvements. Future studies are warranted to improve and further validate the test criteria.

CONCLUSIONS

In the current study, the recently proposed respirator test criteria (Shaffer *et al*, 2014) have been utilized for the first time to evaluate individual features of prototype and commercial filtering facepiece respirators associated with comfort and tolerance issues that are thought to impact user compliance. This study will serve as a baseline endeavor at criteria utilization in the hope of spawning future laboratory and work site studies that are needed to determine the performance of other filtering facepiece models against the proposed criteria.

Acknowledgements

The authors thank Caroline M. Ylitalo, PhD, John M. Sebastian, PhD and Andy Viner, MS, of Company A, Michael Parham, MS of Company B, and Lewis Radonovich Jr., MD, and W. Jon Williams, PhD, of NIOSH, for their manuscript reviews.

Disclaimer

The findings and conclusions of this article are those of the authors and do not necessarily represent the views of the National Institute for Occupational Health and Safety. Mention of a commercial product or trade name does not constitute endorsement by the National Institute for Occupational Safety and Health.

REFERENCES

- Antunano MJ, Baumgardner FW, Chen YT, Constable SH. (1993) Characterization of chemical defense mask breathing resistance tolerances. Air Force Materiel Command, Brooks Air Force Base, Texas, Report AL-TP-1992-0061. Cited November 1, 2011. Available from: <http://www.dtic.mil/cgibin/GetTRDoc?Location=U2&doc=GetTRDoc.pdf&AD=ADA265221>.
- Bergman MS, He X, Joseph ME, Zhuang Z, Heimbuch BK, Shaffer RE, Choe M, and Wander JD. (2015) Correlation of respirator fit measured on human subjects and a static advanced headform. *J. Occup. Environ. Hyg.* 12(3): 163-171.
- Borg G. Psychophysical scaling with applications in physical work and the perception of exertion. (1990) *Scand. J. Work Environ. Health.* 1990; 16 (Suppl 1): 55-58.
- Federal Register / Vol. 74, No. 238 / Monday, December 14, 2009 / Department of Veterans Affairs. Project Better Respiratory Equipment using Advanced Technologies for Healthcare Employees (B.R.E.A.T.H.E.). Cited November 1, 2016. Available from <https://www.gpo.gov/fdsys/pkg/FR-2009-12-14/pdf/E9-29709.pdf>.
- Frank SM, Raja SN, Bulcao C, and Goldstein D. (1999) Relative contribution of core and cutaneous temperatures to thermal comfort, and the autonomic response in humans. *J. Appl. Physiol.* 86(5): 1588-1593.
- Gosch ME, Shaffer RE, Eagan AE, Roberge RJ, Davey VJ, and Radonovich LJ. (2013) B95: a new respirator health care personnel. *Am. J. Infect. Control.* 41(12): 1224-1230.
- Kim H-J, Roberge RJ, Powell JB, Shaffer AR, and Williams WJ. (2012) Measurement accuracy of heart rate and respiratory rate during graded exercise and sustained exercise in the heat using the Zephyr Bioharness™. *Int. J. Sports Med.* 34(6): 497-501.
- National Personal Protective Technology Laboratory (NPPTL) (2012). Standard respirator testing procedures. Cited November 1, 2016. Available: <http://www.cdc.gov/niosh/npptl/stps/APresp.html>
- Occupational Safety and Health Administration (OSHA). Appendix B-1 to § 1910.134: User Seal Check Procedures (Mandatory). Cited November 1, 2016. Available from: https://www.osha.gov/pls/oshaweb/owadisp.show_document?p_table=STANDARDS&p_id=9781.

- Occupational Safety and Health Administration (OSHA). Appendix A to § 1910.134: Fit Testing Procedures (Mandatory). Cited November 1, 2016. Available from: https://www.osha.gov/pls/oshaweb/owadisp.show_document?p_table=STANDARDS&p_id=9780.
- Radonovich Jr, L.E., A. Baig, R.E. Shaffer, R.J. Roberge, A. Levinson, D.W. Doerr, and V. Davey. Better Respiratory Equipment using Advanced Technologies for Healthcare Employees (Project B.R.E.A.T.H.E.). A Report of an Interagency Working Group of the U.S. Federal Government. February 2, 2009. Cited November 1, 2016. Available from: <http://www.publichealth.va.gov/docs/cohic/project-breathe-report-2009.pdf>.
- Roberge RJ, Coca A, Williams WJ, Powell JB, and Palmiero AJ. (2010) Physiological impact of the N95 filtering facepiece respirator on healthcare workers. *Resp. Care*. 55(5): 569-577.
- Roberge RJ, Kim J-H, and Benson S. (2012a) N95 filtering facepiece respirator deadspace temperature and humidity. *J. Occup. Environ. Health*. 9(3): 166-171.
- Roberge RJ, Kim J-H, and Benson SM. (2012b) Absence of consequential changes in physiological, thermal and subjective responses from wearing a surgical mask. *Resp. Physiol. Neurobiol*. 181(1): 29-35.
- Roberge RJ, Kim J-H, Powell JB, Shaffer RE, Ylitalo CM, and Sebastian JM. (2013) Impact of low filter resistances on subjective and physiological responses to filtering facepiece respirators. *PLoS One* 8(12): e84901. doi:10.1371/journal.pone.0084901
- Shaffer RE and Jansen LL. (2015) Selecting models for a respiratory protection program: What can we learn from the scientific literature? *Am. J. Inf. Control*. 43(2): 127-132.
- Shaffer RE, Zhuang Z, Bergman MS, Roberge RJ, and Kim J-H. (2014) Recommended requirements, test methods, and pass/fail criteria for a "B95" respirator for healthcare workers. *J. Int. Soc Resp. Protect*. 31(1): 23-44.
- Smith RM and Rae A. (1977) Thermal comfort of patients in hospital ward areas. *J. Hyg. (London)*. 78(1): 17-26.
- Wilkinson IJ, Pisaniello D, Ahmad J, and Edwards S. (2010) Evaluation of a large-scale quantitative respirator-fit testing program for healthcare workers: survey results. *Infect. Control. Hosp. Epidemiol*. 31(9): 918-925.
- Wizner K, Stradtman L, Novak D, and Shaffer R. (2016) Prevalence of respiratory protective devices in US health care facilities. Implication for emergency Preparedness. *Workplace Health & Safety*. 64(8): 359.
- Zhuang Z, Bradtmiller B and Shaffer RE. (2007) New respirator fit test panels representing the current US civilian work force. *J. Occup. Environ. Hyg*. 4(9): 647-659.
- Zhuang Z, Bergman MS, Lei Z, Niezgoda G, and Shaffer RE. (2017) Recommended test methods and pass/fail criteria for a respirator fit capability Test of half-mask air-purifying respirators. *J. Occup. Environ. Hyg*. (in press).